


Identifier: TC 1-MAN-2.2	Revision: 0	Effective Date: 6/8/16	 QUANTUM N · O · R · T · H · W · E · S · T
Web Page: http://www.qnw.com			
Description: TC 1 Temperature Controller for Quantum Northwest Peltier-Controlled Cuvette Holders			

QNW TC 1 Temperature Controller User's Guide

Quantum Northwest, Inc.
22910 E Appleway Avenue, Suite 4
Liberty Lake, WA 99019-8606

Phone: (509) 624-9290
Fax: (509) 624-9488
Service email: service@qnw.com

www.qnw.com

Temperature-Controlled Cuvette Holders
and Custom Instrumentation

Notices

Copyright

Copyright © 2016 Quantum Northwest, Inc. All rights reserved.

No part of this publication may be reproduced, transmitted, transcribed, stored in a retrieval system, or translated into any language, in any form or by any means, electronic, mechanical, photocopying, recording, or otherwise, without prior written permission from Quantum Northwest, Inc. All copyright, confidential information, patents, design rights and all other intellectual property rights of whatsoever nature contained herein are and shall remain the sole and exclusive property of Quantum Northwest, Inc. The information furnished herein is believed to be accurate and reliable. However, no responsibility is assumed by Quantum Northwest, Inc. for its use, or for any infringements of patents or other rights of third parties resulting from its use.

The Quantum Northwest name, QNW acronym and Quantum Northwest logo are trademarks or registered trademarks of Quantum Northwest, Inc. All other trademarks are the property of their respective owners.

Quantum Northwest products must only be used in the manner described in the Quantum Northwest produce user guides. Any other use may result in damage to the product or personal injury. Quantum Northwest is not responsible for any damages caused, in whole or in part, by improper use of the products, unauthorized alterations, adjustments or modifications to the products, failure to comply with procedures in Quantum Northwest product user guides, or use of the products in violation of applicable laws, rules or regulations. Quantum Northwest shall not be liable for errors or damages on products or performance due to using non-Quantum Northwest genuine parts.

Warranty

The material contained in this document is provided "as is," and is subject to being changed, without notice, in future editions. Further, to the

maximum extent permitted by applicable law, Quantum Northwest disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Quantum Northwest shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Quantum Northwest and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Safety Notices

CAUTION

A CAUTION notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a CAUTION notice until the indicated conditions are fully understood and met.

WARNING

A WARNING notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood and met.

Table of Contents

1.	General Information	5
1.1.	User Documentation	5
1.2.	Other Precautions	5
1.3.	Warning and Other Symbols	5
1.4.	CE Compliance	6
	Declaration of Conformity to EU Directive 2011/65/EU (RoHS)	8
1.5.	Introduction: QNW TC 1 Temperature Controller	9
1.6.	Measurement Category	9
1.7.	Pollution Degree	9
1.8.	Overvoltage Category	9
1.9.	Environmental Conditions	9
1.10.	Electrical Specifications	10
1.11.	Computer Requirements	10
2.	Installing, Starting and Stopping the QNW TC 1 Temperature Controller	11
2.1.	Installing the QNW TC 1 Temperature Controller	11
2.2.	Starting the QNW TC 1 Temperature Controller	13
2.3.	Stopping the QNW TC 1 Temperature Controller	14
3.	Step-by-Step Instructions for Using the QNW TC 1 Temperature Controller	15
3.1.	Instrument Overview	15
3.2.	Detailed Instructions	15
3.3.	External Computer Control	17
3.4.	Troubleshooting	18
3.5.	Error Messages	18
4.	Maintenance and Spare Parts	20
4.1.	Maintenance	20
4.2.	Spare Parts	20
5.	Serial Communications for the TC 1 Temperature Controller	21
5.1.	Sample Holder ID Number	22
5.2.	Controller Firmware Version	22
5.3.	Stirrer	22
5.4.	Temperature Control	23

5.5.	Target Temperature	23
5.6.	Instrument Status	24
5.7.	Current Sample Holder Temperature.....	24
5.8.	Error Reporting.....	24
5.9.	Probe Status and Temperature.....	25
5.10.	Temperature Ramping	26
5.11.	Heat Exchanger Temperature	27
5.12.	Cell Changing	28
5.13.	Reference Cuvette	28
5.14.	Front Panel Controls	29
6.	Elements of the T-App Software Help File	30
6.1.	Application Startup Window	30
6.2.	Application Control Window	31
6.3.	Application Menus.....	32
6.4.	File Menu	32
6.5.	Tools Menu	32
6.6.	Temperature Control.....	32
6.7.	Temperature Ramping	33
6.8.	Stirrer Control.....	33
6.9.	Show Script Panel.....	33
6.10.	Command Text Box.....	35
6.11.	Plot Menu	35
6.12.	Data Shown	36
6.13.	Single Sample Status Panel.....	36
6.14.	Temperature Control Status	37
6.15.	Heat Exchanger Temperature	37
6.16.	Target Temperature	38
6.17.	Magnetic Stirrer Controls.....	38
6.18.	External Probe Status	38
6.19.	Plot Axis Label Manipulation	39
6.20.	Plot of Data Collected.....	39
6.21.	Scripts: Program Script Commands	40

1. General Information

Thank you for purchasing a Quantum Northwest **TC 1 Temperature Controller** and associated Peltier-controlled cuvette holder. We want you to enjoy many years of faithful service from your instrument. If you have any questions, feel free to contact us directly through our website, www.qnw.com/contact-us/, or by email at service@qnw.com.

Your Quantum Northwest temperature controller and associated cuvette holder have been carefully designed so that when used properly, you have an accurate, fast, flexible and safe temperature control system.

Information about safety practices appears throughout the documentation provided with your instrument and accessories to help you safely operate the instrument and accessories. Before using the instrument or accessories, you must thoroughly read these safety practices. **ALWAYS** operate the instrument and accessories in accordance with these safety practices.

1.1. User Documentation

You have been provided with the following documentation to help you set up and operate your Quantum Northwest **TC 1 Temperature Controller**:

- This manual, with safety practices and hazards information, instructions for installing and maintaining the components of the **QNW TC 1 Temperature Controller**, and troubleshooting information.
- The **T-App** Help file, located on the **T-App** CD, if purchased (Quantum Northwest>**T-App** Help).

1.2. Other Precautions

Do not block any ventilation grills present on the **QNW TC 1 Temperature Controller**.




Use of the **QNW TC 1 Temperature Controller** and accessories may involve materials, solvents and solutions that are flammable, corrosive, toxic or otherwise hazardous. Careless, improper, or unskilled use of such materials, solvents and solutions can create explosion hazards, fire hazards, toxicity and other hazards which can result in death, serious personal injury, and damage to equipment and property.

ALWAYS ensure that laboratory safety practices governing the use, handling and disposal of such materials are strictly observed. These safety practices should include the wearing of appropriate safety clothing and safety glasses.

1.3. Warning and Other Symbols

The following is a list of symbols that appear in conjunction with warnings on the QNW cuvette holder and accessories. The hazard they describe is also shown.

A triangular symbol indicates a warning. The meanings of the symbols that may appear alongside warnings in the documentation or on the instrument itself are as follows.

	When used on a warning label, the user is in danger of an electrical shock.
	When used on warning labels attached to the instrument, refer to the relevant operation or service manual for the correct procedure referred to by that warning label.
	When attached to the rear of the instrument, indicates that the product complies with the requirements of one or more EU directives.

1.4. CE Compliance

Your **QNW TC 1 Temperature Controller** has been designed to comply with the requirements of the Electromagnetic Compatibility (EMC) Directive and the Low Voltage (electrical safety) Directive (commonly referred to as the LVD) of the European Union. Quantum Northwest has confirmed that each product complies with the relevant Directives by testing a prototype against the prescribed EN (European Norm) standards.

Proof that a product complies with these directives is indicated by:

- the CE Marking appearing on the rear of the product, and
- the following Declaration of Conformity. The Declaration of Conformity is the legal declaration by Quantum Northwest that the product complies with the directives listed above, and shows the EN standards to which the product was tested to demonstrate compliance.

EU Declaration of Conformity

We: Quantum Northwest, Inc.
 Of: 22910 E Appleway Ave, Suite 4
 Liberty Lake, WA 99019-8606
 USA

Hereby declare under our sole responsibility that the following devices:

QNW MODEL (CONTROLLER MODEL)
BATH 10
CD 250™ (TC 1 Controller)
Luma 40™ (TC 1 Controller)
Peltier-based cuvette holder (TC 1 Controller)
qpod® (TC 1 Controller)
t2 Sport™ (TC 1 Controller)
t2x2 Sport™ (TC 1 Controller)
Turret 6™ (TC 1 Controller)
Versa 20™ (TC 1 Controller)

Comply with the essential requirements of the following applicable European Directives:

Electromagnetic Compatibility (EMC) Directive 2004/108/EC,
Low-Voltage Directive 2014/35/EU,
RoHS Directive 2011/65/EU.

Conformity is assessed in accordance to the following standards:

EMC Directive:	Emissions	EN 61326-1 (2013) Group 1 Class A
	Immunity	EN 61326-1 (2013) Group 1 Class A
Low Voltage Directive:	Safety	EN 61010-1

I hereby declare that the equipment models named above have been assessed and found to comply with the relevant sections of the above-referenced specifications. The units comply with the relevant requirements of the applicable Legislation, and I am the person authorized to compile the technical documentation.

Signed by: 

Name: Enoch W. Small, Ph.D.
 Position: President, Quantum Northwest, Inc.
 Date: May 24, 2016
 Location: Liberty Lake, Washington, USA

Declaration of Conformity to EU Directive 2011/65/EU (RoHS)

We: Quantum Northwest, Inc.
 Of: 22910 E Appleway Ave, Suite 4
 Liberty Lake, WA 99019-8606
 USA

Declare under our sole responsibility that our products indicated below are in fully compliance with EU Directive 2011/65/EU, with respect to the following substances:

1. Lead (Pb),
2. Mercury (Hg),
3. Cadmium (Cd),
4. Hexavalent chromium (Cr (VI)),
5. Polybrominated biphenyls (PBB),
6. Polybrominated diphenyl ethers (PBDE), Deca-BDE included.

QNW MODEL (CONTROLLER MODEL)
BATH 10
CD 250™ (TC 1 Controller)
Luma 40™ (TC 1 Controller)
Peltier-based cuvette holder (TC 1 Controller)
qpod® (TC 1 Controller)
qpod® 2e (Integrated Control)
t2 Sport™ (TC 1 Controller)
t2x2 Sport™ (TC 1 Controller)
TLC 50™ Legacy (TC 125 Controller)
Turret 400™ (TC 425 Controller)
Turret 6™ (TC 1 Controller)
Versa 20™ (TC 1 Controller)

The following statement regarding conformity of our products to the EU Directive 2011/65/EU (RoHS) is based on our actual research and inquiries with our suppliers:

All products sold under the above model numbers previously, currently and in future are/will be fully compliant with EU Directive 2011/65/EU and without exemption.


Signed by: 
 Name: Enoch W. Small, Ph.D.
 Position: President, Quantum Northwest, Inc.
 Date: April 11, 2016
 Location: Liberty Lake, Washington, USA



Figure 1. TC 1 Temperature Controller

1.5. Introduction: QNW TC 1 Temperature Controller

Intended Use

The **QNW TC 1 Temperature Controller** is used with an associated Peltier-controlled cuvette holder to control sample temperature during optical measurements. Temperature control is achieved using the **TC 1 Temperature Controller** alone or along with a computer interfaced to the **TC 1 Temperature Controller** box, using the **QNW T-App** software provided.

1.6. Measurement Category

The measurement category is IEC 61010:1. Do not to use this equipment for measurements within measurement categories II, III and IV.

1.7. Pollution Degree

The pollution degree is IEC61010:2. Pollution degree “2” applies to a normal indoor atmosphere.

1.8. Overvoltage Category

The overvoltage category (installation category) is CAT II. See IEC 664 & IEC 61010.

1.9. Environmental Conditions

The area should have a dust-free atmosphere with minimal drafts, vibrations, and corrosive fumes. For optimum performance, the ambient air is recommended to be between 20 and 25 °C, but can be from 5 to 40

°C. Relative humidity should be less than 80%. The instrument is designed for operation at 2,000 m elevation or less.

To avoid damage through spillage of solutions and samples being analyzed, the worktops should be covered with a material that is corrosion resistant and impervious to liquids.

Allow at least two inches of space on both sides, and six inches at the rear of the instruments to permit free air circulation. The power cord and all other connections are located at the rear of the instruments. The power switch is located on the front panel.

1.10. Electrical Specifications

Mains Supply

A standard power cable is provided based on the user's country requirements (18 AWG/115 V AC US/Canada; 1.0 mm/220 V AC international). The required supply voltage is 100-240 V AC (frequency 50 or 60 Hz). The mains voltage fluctuations must not exceed 10% of the nominal voltage.

The installation of electrical power supplies must comply with the rules and/or regulations imposed by local authorities responsible for the supply of electrical energy to the workplace.

All power supplies for the **TC 1 Temperature Controller** must be single-phase, AC voltage, three-wire system (active, neutral, earth) and should be terminated at an appropriate power outlet receptacle that is within reach of the power cord. For safety reasons, a separate power outlet receptacle should be provided for each unit in the system. The use of extension cords or outlet adaptors is not recommended.

WARNING



Shock Hazard - Danger of electrocution. Good electrical grounding is essential to avoid potentially serious shock hazards. A 3-wire outlet with ground connection must be provided for the **TC 1 Temperature Controller**. Make certain that power outlets are earth-grounded at the grounding pin.

CAUTION



Caution - Never connect or disconnect any cables while the **TC 1 Temperature Controller** is switched on. Damage to the printed circuit boards may occur.

1.11. Computer Requirements

The **T-App** temperature application software used for external control of Quantum Northwest temperature controllers requires a personal computer using a Microsoft Windows™ operating system (XP, Vista, 7, or 8). The **T-App** software is normally provided on a CD, requiring a CD player be part of the computer. However, **T-App** may also be downloaded *via* the Internet from a QNW ftp site. Contact QNW should you wish to obtain the application by Internet.

2. Installing, Starting and Stopping the QNW TC 1 Temperature Controller

2.1. Installing the QNW TC 1 Temperature Controller

Unpacking

After accepting delivery, take the equipment to the installation site. Quantum Northwest instruments are inherently robust, and the packaging is designed to prevent internal damage. However, the contents form part of a precision measuring system and all packages should be handled with care. In transit, sharp jolts must be avoided and the packages should not be inverted or tilted unnecessarily. Markings on the shipping cartons generally indicate which side of the package should be kept on top.

Unpacking the equipment is your responsibility. As the packages are opened, ensure that you have received everything you ordered. If there are any discrepancies, notify the supplier. If any items are found to be damaged, immediately notify the carrier and supplier.

You should have received:

- 1x **TC 1 Temperature Controller**
- 1x power cable
- 1x USB cable
- 1x CD containing **T-App** software (if ordered)
- This manual

Note: Any additional items ordered are not listed here.

Any differences from the original order should be referred immediately to your Quantum Northwest sales office. Do not discard any packaging components or filler materials.

Installing the QNW T-App Software (if ordered)

Follow the instructions provided with the software to complete the installation. The instructions are found in a text file titled "Instructions.txt" within the **T-App** folder. Those instructions will be similar to the following:

Instruction for installation of T-App

1. Copy the "T-App" folder to a convenient location on the computer to be used.
2. Run "CDM v2.12.12 WHQL Certified.exe" to the install drivers needed for serial communications between the QNW TC 1 temperature controller and the T-App application.
3. Connect the TC 1 to the computer to be used with the USB cable provided with the TC 1 (or any USB cable with the correct plug ends). Turn on the TC 1 power. Installation of the drivers for the USB connection should begin immediately and proceed automatically.
4. Run "T-App.exe". For further information select the Help>(T-App Help) menu item.

Note: It is important to install software before connecting the **TC 1** to your computer. Software installation installs a driver. If the **TC 1** Temperature Controller is connected to the computer and turned on before software installation, then the computer may automatically load an incorrect driver that may be difficult to remove.

Connect the **TC 1** to your computer using the USB or RS 232 cable provided. The New Hardware installation process will begin automatically and take a few moments to finish.

Installing the QNW TC 1 Temperature Controller Hardware



Figure 2. Example of a user-supplied submersible pump, bucket, electrical cable, and vinyl tubing to be used with the **TC 1 Temperature Controller** and associated Peltier-controlled cuvette holder

1. Install the Peltier-controlled cuvette holder into the destination spectrometer.

Note: The **TC 1 Temperature Controller** and its associated Peltier-controlled cuvette holder are supplied completely assembled. It is necessary only to make the appropriate connections.

2. Position the bucket to within 75 cm (30 in) above or below the spectrometer.

CAUTION



Caution – Position the bucket in a location that minimizes the risk of spilling the liquid contents.

3. OPTIONAL - If a gas such as nitrogen is to be used as dry gas for minimizing condensation at low temperatures, secure the tank to a solid object in the vicinity of the spectrometer.

CAUTION



Caution – When using dry gas do prevent condensation on the cuvette, ensure that the working area is adequately ventilated.

4. Cut the vinyl tubing into two lengths (three if you will be using the dry gas purge). Attach one length of tubing from the hose barb of the submersible pump, shown in Fig. 2, to either one of the water hose barbs on the Peltier-controlled cuvette holder, labeled "water access for thermoelectric cooler." Attach another piece of tubing from the second cuvette holder water hose barb to the bucket to serve as a drain tube. Add water to the bucket and lower the pump into the bucket, making sure that the pump is below the water line. Using the electrical cable from the pump, connect the pump to the socket on the back of the **TC 1 Temperature Controller**. The pump will start automatically when the **TC 1** power is turned on later.
5. OPTIONAL - If you plan to work at low temperatures, connect a source of dry gas (typically nitrogen) to release upon the cuvette windows. Attach a length of tubing with 1/8-inch inside diameter to the small dry gas hose barb on the base of the Peltier-controlled cuvette holder. Set the dry gas flow rate to about 50 cc/min.
6. Connect the electrical cable from the cuvette holder to the 15-pin, VGA-type connector labeled "Sample" on the back of the **TC 1 Temperature Controller**.
7. Plug the power cable into the back of the **TC 1** Temperature Controller and into a wall socket.

2.2. Starting the QNW TC 1 Temperature Controller

1. Place the magnetic stir bar in a standard 1 x 1-cm square cuvette. Insert a liquid sample into the cuvette.
2. Place the cuvette in the cuvette holder.
3. Place the plastic lid on the cuvette holder.
4. OPTIONAL - If you wish to monitor the temperature inside the cuvette, plug a standard Series 400 or Series 500 thermistor probe (not provided) into the 1/4 - inch phone jack labeled "probe" in the back panel of the **TC 1**. Place the end of the probe in a region of the sample in the cuvette where it will not occlude the spectrometer light beam.
5. Turn on the **TC 1** controller using the switch on the back panel. On the **TC 1** front panel, press once on the right-arrow button and once on the SET button. These two steps turn on the pump by turning on the temperature control. (The pump runs only when the temperature control is turned on).
6. Check for water leaks.
7. Check that water is flowing from the free tubing end. If it does not, lift the pump off the bottom of the bucket, but not out of the water, and drop it. This usually will get the flow started. If it does not after several tries, lift the pump up close to the water surface and let a small drop of liquid detergent land in the water close to the pump intake.
8. For now, turn the **TC 1** power off.




Figure 3. TC 1 Menu Button

-
9. Use the left or right arrows on the Menu Button, shown in Fig. 3, to cycle through the four pages of options. These options are described in detail in Section 2.3 below. To turn on stirring and set a temperature, go to the Set Stirring page, use the up and down arrows to set a speed or let it default to 1200 rpm, and press SET to start the stirrer. Then go to the Set Target Temperature page and use the up and down arrows to set a target temperature, for example, 37.0 °C. Press SET to initiate temperature control. The temperature will then change and stabilize.

2.3. Stopping the QNW TC 1 Temperature Controller

1. While temperature control is running and the Start page is being displayed, press SET (Fig. 3) to turn off temperature control. To turn off stirring, go to the Set Stirrer page and press SET again.
2. After measurements are completed, turn off power on the back of the **TC 1** controller. This action should power off both the **TC 1** and the **BATH 10** pump, if used. If you are using a circulator pump other than the **BATH 10**, power it off at this time.

3. Step-by-Step Instructions for Using the QNW TC 1 Temperature Controller

3.1. Instrument Overview

The **TC 1 Temperature Controller** may be used in a stand-alone manner together with its associated Peltier-controlled cuvette holder, or under PC control using QNW's **T-App** software or other vendors' software.

T-App is a program that permits external computer control of the **TC 1 Temperature Controller**. The temperature of the sample holder and a temperature sensed by an external probe may be plotted vs time. Simple text scripts may be used to automate multiple operations. **T-App** is provided with a USB cable to connect your computer to the **TC 1**. Full instructions for using **T-App** may be found in the associated Help file.

3.2. Detailed Instructions

Using the Menu Button control arrows




Figure 4. TC 1 Menu Button control arrows

Refer to Fig. 7 for the appearance of the **TC 1** Menu button. Use the right and left arrows to cycle through four pages:

– **Display** – **Set Target Temperature** – **Set Stirring** – **Set Ramping** –

Display

Holder =	37.0 °C
Target =	37.0 °C
Probe =	36.9 °C
Ramp On	Stir On

Figure 5. Example **TC 1** Display page for a single cuvette holder and external probe, with temperature

Display: This main page shows the actual cuvette *Holder* temperature, the *Target* temperature and a *Probe* temperature (if a probe is present). The page also shows whether the *Ramp* and *Stir* functions are on or off. After a few seconds of inactivity, all other pages will revert back to the **Display** page.

When seeking a temperature, the green LED on the front panel will flash slowly. When the cuvette holder has locked onto a new target temperature, the green LED will remain lit. A rapidly flashing red LED usually means a loose electrical cable or inadequate water flow for the Peltier unit.

Note: Pressing the **SET** button while the **Display** page is active results in no action. To start or stop controller functions, access the four other pages, **Set Target Temperature – Set Stirring – Set Ramping – Set Position**.

Set Target Temperature

Set Target Temp.	
Target =	37.0 °C
Current=	23.6 °C Off

Figure 6. **TC 1** Set Target Temperature page

Set Target Temperature: To set the *Target* temperature, use the up and down arrows. Press **SET** to retain this new *Target* and initiate temperature control. The green LED light will begin flashing slowly as the device seeks the targeted temperature, and the page window will show the *Current* to be *ON*.

Discontinue Temperature Control: Press **SET** while the green LED light is lit or slowly flashing to discontinue temperature control. The LED light will turn off completely, and the page window will show the *Current* to be *Off*.

Set Stirring

Set Stirring	
Stir speed =	1200 rpm
Current =	Off

Figure 7. **TC 1** Set Stirring page

Set Stirring: To turn on magnetic *Stirring*, use the up and down arrows to choose an approximate stirring speed between 200 and 2000 rpm. Press **SET** to set the speed and initiate stirring. The page will update to show the *Current* value to be the same as *Stir speed*.

Discontinue Stirring: Press **SET** while the *Stirring* is on (shown by nonzero *Current* rpm) to turn the stirrer off. The page will update to show the *Current* speed to be *Off*.

Set Ramping

Set Ramping

Ramp = 0.55°/min

Current = Off

Figure 8. TC 1 Set Ramping page

Set Ramping: To perform a temperature ramp, set the *Ramp* rate using the up and down arrows, and press **SET**. The page will update to show the *Current* rate in °/min to be the same as the *Ramp* rate. With ramping set, turning on temperature control will generate a linear ramp to the target temperature. When the sample holder reaches the target temperature, no further temperature change occurs, although the *Current* rate will remain as set.

The fastest possible ramp is determined by how fast the cuvette holder is able to reach the target temperature without ramping. Attempting to ramp too quickly, especially at high and low temperature extremes, will result in a nonlinear ramp. The slowest ramp that may be set on this page is 0.09 °C/minute. (If needed, much slower ramps may be set through software commands.)

Discontinue Ramping: Press **SET** while the *Ramp* is on (shown by nonzero *Current* °/min) to turn the *Ramp* off. The page will update to show the *Current* rate to be *Off*.

3.3. External Computer Control

All functions may be accessed either through a Serial (RS 232) or a USB located on the back of **TC 1 Temperature Controller**. You may write your own program or purchase the QNW application program **T-App**. **T-App** will plot temperatures of the probe, cuvette tower, and even the Peltier element heat exchanger vs time. It will also permit you to set up temperature ramps. If you wish to do your own programming, please see the Appendix for communication instructions and the set of text commands that may be used and responses to the commands.

3.4. Troubleshooting

If the submersible pump has operated without being immersed in water, it may not prime properly. If this occurs, place one drop of detergent on the pump intake and try again.

If the **TC 1** display indicates that there is inadequate coolant, it may be that the circulating water is too warm. Add ice to the water to cool it down.

If a circulating bath is used as the water source, it may not be possible to maintain a flow rate of 200 mL per minute. While this condition will limit the lowest temperature attainable, it may be partially compensated by using colder circulating liquid. Below about -25 °C, 30% methanol in water is generally recommended as the coolant.

If the set temperature is near room temperature and too close to the temperature of the cooling water, the sample holder temperature may oscillate around the set temperature. In this case, changing the temperature of the cooling water (for example by adding ice) should stop the oscillation. Cooler circulating water can result in improved performance of the cuvette holder at low temperatures, whereas warmer water can improve performance at very high temperatures.

If a cuvette shorter than 44 mm is used, it may be difficult to remove. Use a taller cuvette instead.

3.5. Error Messages

When errors occur, the line 1 of the **TC 1** Temperature Controller display presents an error code. Line 3 of the display identifies the error and line 4 of the display presents possible solutions. The most common events that cause errors to be displayed are loose cables or inadequate coolant flow. For errors not easily solved, contact QNW at www.qnw.com/contact-us/ or service@qnw.com.

Specific error messages include the following.

5 – cell out of range

warnings: loose cable, sensor failure

The temperature controller is not receiving a reasonable response from the sensor on the cuvette tower. Either the sensor has failed or a cable is not making a good connection.

6 – cell out of range

warnings: loose cable, check cable

The temperature controller is not receiving reasonable responses from either the cell tower or heat exchanger sensors. Since it is very unlikely for both to fail, probably a cable is loose.

7 – heat exchanger error

warnings: loose cable, sensor failure

The temperature controller is not receiving a reasonable response from the sensor on the heat exchanger. Either the sensor has failed or a cable is not making a good connection.

8 - inadequate coolant

warnings: inadequate coolant, water temperature

The sensor on the heat exchanger is reading a temperature above 60 °C. Temperature control has been shut down to prevent damage to the Peltier element. Either the water was too warm or the rate of flow was inadequate to draw sufficient heat from the heat exchanger.

9 - Invalid command

The controller has been sent an invalid text command.

4. Maintenance and Spare Parts

4.1. Maintenance

The **TC 1 Temperature Controller** requires very little routine maintenance on the part of the user. For routine cleaning of exposed surfaces, use only a cloth dampened with water or diluted detergent. Do not use organic or abrasive solvents.

CAUTION



Caution – Any action which makes it necessary to open the **TC 1 Temperature Controller** unit must be executed only by QNW technicians or authorized personnel.

The water hoses and the submersible pump should be inspected prior to each usage to ensure that they are intact. Replace the nylon tubing when it becomes discolored or cracked. Periodically replace the water stored in the bucket to minimize the growth of microorganisms. When not in use, the pump, bucket, and tubing may be stored dry.

4.2. Spare Parts

Vinyl tubing and magnetic stir bars may be obtained from a variety of commercial vendors. For electrical cables to be used with **TC 1 Temperature Controller** and submersible pumps, please contact QNW at service@gnw.com.

5. Serial Communications for the TC 1 Temperature Controller

2/5/2016

Note: This document provides the serial communications protocols for version 2.2 of the firmware on the TC 1 family of controllers.

TC 1/Single – for single cuvette holders such as the t2 Sport, Versa 20, Luma 40, CD 250 and Flash 300;

TC 1/Dual – for dual independent control for sample and reference cuvettes of the t2x2 Sport and Versa Dual;

TC 1/Multi – for multi-position sample changers such as the qCHANGER 6 and the Turret 6.

The version number and the ID (see below) are shown briefly on the display when the temperature controller is turned on. This initial display can be repeated by pressing the down arrow on the front panel of the controller.

All functions of the temperature controller can be managed from a computer, using the command set described below. If you purchased your unit as a component of a spectrometer from certain manufacturers, this feature may be implemented through traditional RS232 serial connectors on the computer, or the spectrometer and on the controller. In this case they will be connected by a standard 15-pin serial extension cable (male connector on one end and female on the other). No driver installation should be needed. Otherwise the serial linkage will be established through a USB connection between the computer and the controller. In this case the controller includes electronics which convert the USB connection to a serial communications port. However, for the port to be available to programs on the computer, it will be necessary to load driver software.

CAUTION



Caution - It is important that the driver software be loaded before connecting a USB cable between the controller and the computer. Contact Quantum Northwest for further information.

Quantum Northwest offers the **T-App** control program written specifically for control of all functions of the temperature controller and to track temperature as a function of time. The resulting data can be saved to a text file (two columns, time and temperature, tab delimited) or copied and pasted directly into another program (such as Microsoft Excel). If you expect to create software or firmware that communicates with a **TC 1** controller, **T-App** may prove extremely useful for preliminary testing of controller commands.

In programming for the **TC 1** controller, one must adhere to the conventional notation: 8/N/1.

Baud:	19200
Data Bits:	8
Parity:	None
Stop Bit:	1
Flow Control:	None

For many of the commands listed below, the controller returns information in response to the command. All commands and responses are delineated by left and right square brackets ([]). Any text sent to the controller not enclosed within brackets will be ignored. In this document, an ellipsis (...) is used to distinguish responses from commands.

Note: The format of the commands is shown below.

[command]	Purpose of the command (sent to the controller).
... [response]	Meaning of the response (received from the controller).

5.1. Sample Holder ID Number

[F1 ID ?]	What is the ID number of the sample holder being controlled?
... [F1 ID 14]	ID is 14.

Assigned Identities:

ID = 00	specialty sample holder (see command class 14)
14	single
24	dual
34	turret or linear multi-sample holder

5.2. Controller Firmware Version

[F1 VN ?]	What is the <u>V</u> ersion <u>N</u> umber of the controller firmware?
... [F1 VN 2.22]	The controller firmware version number is 2.22.

5.3. Stirrer

[F1 MS ?]	What is the <u>M</u> aximum stirrer <u>S</u> peed?
... [F1 MS 2500]	The maximum stirrer speed permitted is 2500 rpm.
[F1 LS ?]	What is the <u>L</u> owest stirrer <u>S</u> peed?
... [F1 MS 300]	The lowest stirrer speed permitted is 300 rpm.
[F1 SS S 1000]	Set <u>S</u> tirrer <u>S</u> peed to 1000 rpm <u>and turn stirring on</u> .
[F1 SS S 0]	Turn stirrer off (does not change the speed setting).
[F1 SS +]	Turn stirrer on and set it to the most recent non-zero stirrer speed setting.
[F1 SS -]	Turn stirrer off (this does not change the speed setting)..
[F1 SS ?]	What is the current stirrer speed setting? Depending on the R+/R- state as specified below, the stirrer status ('+' /on or '-' /off) may also be reported.
... [F1 SS 1000]	Stirrer speed setting is 1000 rpm.

... [F1 SS -]	Stirrer status is off.
[F1 SS R-]	Turn off automatic R eporting of the stirrer speed and status when changed by a command.
[F1 SS R+]	Turn on automatic reporting of stirrer changes. The power on default setting is to not report the speed or the status when changed. If you then send [F1 SS R+], only the stirring speed will be reported when changed. If you then send [F1 SS R+], both the speed and the status will be reported when changed. If you then send [F1 SS R-], neither the speed nor the status will be reported when changed (back to the power on default setting). Note that the [F1 SS ?] will always result in a stirring speed response. The status response will also be sent if it has been enabled as described above.

5.4. Temperature Control

[F1 TC +]	Turn T emperature C ontrol on.
[F1 TC -]	Turn temperature control off.
[F1 TC ?]	What is the current status of temperature control?
... [F1 TC -]	Temperature control status is ' - ' /off.
[F1 TC R+]	R eport temperature control status when changed by a command.
[F1 TC R-]	Do not report temperature control status when changed

5.5. Target Temperature

[F1 MT ?]	What is the M aximum T arget temperature setting allowed for this holder?
... [F1 MT 105]	The maximum target temperature allowed is 105 °C.
[F1 LT ?]	What is the L owest T arget temperature setting allowed for this holder?
... [F1 LT -30]	The lowest target temperature allowed is -30 °C
[F1 TT S 23.10]	Set the T arget T emperature to 23.10 °C. Note: The TC 1 does not turn temperature control on when TT is received (as it does when the front panel controls are used to set the TT).
[F1 TT ?]	What is the current target temperature?
... [F1 TT 71.32]	Target temperature is 71.32 °C.
[F1 TT +] or [F1 TT R+]	Turn on automatic reporting of the target temperature when changed by a command.
[F1 TT -] or [F1 TT R-]	Turn off automatic reporting of target temperature changes.

5.6. Instrument Status

[F1 IS ?]	What is the current <u>I</u> nstrument <u>S</u> tatus?
... [F1 IS 0-+S]	Response is four parameters (or five, see below): number of unreported errors is 0 (0 or 1); stirrer is off (+ is on, - is off); temperature control is on (+ is on, - is off); current sample holder temperature is stable (S is stable, C is changing).
[F1 IS +] or [F1 IS R+]	Automatically report instrument status whenever it changes (e.g., when the sample holder temperature goes from changing to stable).
[F1 IS -] or [F1 IS R-]	Stop automatic reports of instrument status when it changes (the power-on default).
[F1 IS E+]	Include the ramp status as a fifth parameter in the instrument status response.
... [F1 IS 0-+SW]	The ramp status will be one of three states represented by the characters ' - ' (minus), ' + ' (plus) or ' W '. See command class 10 for details.
[F1 IS E-]	Stop including the ramp status in the instrument status response (power-on default).

5.7. Current Sample Holder Temperature

[F1 CT ?]	What is the <u>C</u> urrent <u>T</u> emperature of the sample holder?
... [F1 CT 22.84]	The current temperature is 22.84 °C.
[F1 CT +3]	Periodically report current temperature every 3 seconds.
[F1 CT -]	Stop periodic current temperature reports.
[F1 CT +]	Restart periodic probe temperature reports using the current interval (the power-on default is a 3-second interval).
[F1 CT R+]	R eport status of the sample holder temperature when it changes.
... [F1 CT C]	The status of CT is changing.
... [F1 CT S]	The status of CT is stable (CT has stayed within +/-0.05 °C of the target temperature for at least 1 minute.
[F1 CT R-]	Stop reporting changes in the status of CT (power-on default).

5.8. Error Reporting

[F1 ER ?]	Report the current error (ER ror).
... [F1 ER -1]	No current error.
... [F1 ER 05]	Cell T out of range (Loose cable? Sensor failure?).
... [F1 ER 06]	Cell and heat exchanger T out of range (Loose cable?).
... [F1 ER 07]	Heat exchanger T out of range (Loose cable? Sensor failure?).

... [F1 ER 08] Inadequate coolant (check flow). Temperature control has shut down.
 ... [F1 ER 09<<bad command>>] Syntax error on a preceding command where "bad command" is the text of the command that caused the syntax error response.
 [F1 ER +] Automatically report errors when they occur.
 [F1 ER -] Stop automatic error reports (this setting have no effect on ER 9 responses).

5.9. Probe Status and Temperature

[F1 PS ?] What is the Status of the TC 1 external Probe connector?
 ... [F1 PR +] A series 400 thermistor probe is connected.
 ... [F1 PR -] No probe is connected.
 [F1 PS +] Enable probe status to be Reported automatically when a probe is installed or removed.
 or [F1 PS R+]
 [F1 PS -] Disable automatic sending of probe status (power-on default).
 or [F1 PS R-]

[F1 PT ?] What is the current Probe Temperature?
 [F1 PT +3] Periodically report the probe temperature every 3 seconds (integers only, the '+' is required).
 ... [F1 PT 22.37] The current probe temperature is 22.37 °C.
 ... [F1 PT NA] Probe temperature is not available.
 [F1 PT -] Stop periodic probe temperature reports. The interval is retained.
 [F1 PT +] Restart periodic probe temperature reports using the current interval (the power-on default is 3 seconds).

[F1 PA S 0.5] Set the temperature interval for Automatic reporting of the Probe temperature to 0.5 °C during a ramp. (Increment must be positive, without sign in tenths between 0.1 and 9.9 °C, and will work for ramps going up or down.)
 Note: this command does not start automatic reporting, it just sets the interval to be used when it is turned on.

[F1 PA ?] What is the current temperature interval for automatic reporting?
 [F1 PA +] Start automatic reporting of probe temperature at temperature intervals (set by the command above).
 [F1 PA -] Stop automatic reporting of probe temperature every temperature increment.

[F1 PX +] Change probe temperature returned to a precision of 0.01 °C.
 [F1 PX -] Change probe temperature returned to a precision of 0.1 °C.

Note: Since probe temperatures are always reported to 0.01 °C by the TC 1 controller, these commands have no effect. They are included to maintain compatibility with customer software / firmware based

on the older TC 125, TC 225 and TC 425 temperature controllers (i.e., they will not cause an [F1 ER 9 <<....>>] report – see command class 8).

Note: Except for [F1 PS ?] and [F1 PS R(+/-)], any probe related command issued when a probe is not connected to the TC 1 will result in the warning response [F1 NOPROBE].

5.10. Temperature Ramping

[F1 RR S 0.50] Sets the ramp rate to 0.50 °C/minute. This command also sets the ramp status to 'W' / waiting mode (see below). With the exception of 0 the ramp rate must be in the range 0.01 to 10; otherwise an ER9 reply will be returned by the controller, the ramp rate will be set to the nearest allowed value, and the controller will send a second response specifying the ramp rate that was set.

[F1 RR S 0]

or [F1 RR -] Sets ramp status to ' - ' /off (but does not change the current ramp rate setting). If either command is sent during a temperature ramp, the controller will then drive at maximum to the last set target temperature.

[F1 RR +]

Set ramp status to 'W' / waiting mode. In waiting mode the TC 1 is waiting for a new target temperature to be set (see command class 5). When the TT command is received, if temperature control is on (see command class 4) the controller begins ramping the temperature from the current value to the new TT at the previously specified ramp rate. If temperature control was off when the TT command is received, the ramping process does not begin until it is turned on. In either case the ramping state is ' + ' /on until the ramp is completed or cancelled.

If this command is sent while a ramp is in progress, the ramp status will change to 'W' / waiting and the controller will drive at maximum to the last target temperature set.

[F1 RR ?]

What is the current ramp rate setting? Depending on the R+/R- state as specified below, the ramp status ('W' /waiting, ' + ' /on or ' - ' /off) may also be reported.

... [F1 RR 1.00]

Current ramp rate is 1.00 °C/minute.

... [F1 RR W]

Ramp status is waiting.

[F1 RR R-]

Turn off automatic reporting of the rate and ramp status when changed by command or by use of the front panel.

[F1 RR R+]

Turn on automatic reporting of ramp changes. Automatic reports will be sent by the controller when the ramp rate or status is changed by command.

The power on default setting is to not report the ramp rate or the ramp state when changed.

If you then send [F1 RR R+], only the ramp rate will be reported when changed.

If you then send [F1 RR R+], both the rate and the state will be reported when changed.

If you then send [F1 RR R-], neither the rate nor the state will be reported when changed (back to the power on default setting).

Note that the [F1 RR ?] will always result in a ramp rate response. The ramp status response will also be sent if it has been enabled as described above.

The following seven commands, are accepted by and work with the TC 1. They are provided mainly to maintain compatibility with control software/firmware developed for use with the older TC 125/225/425 family of temperature controllers.

There are no corresponding reference commands (see command class 13) for the TL commands.

[F1 TL +] Ramp the sample and reference identically.

[F1 TL -] or.

[F1 TL 0] Ramp the sample and reference independently (the power on default).

[F1 RS S #], [F1 RS ?] Set or query the RS parameter (# is a positive integer).

[F1 RT S #], [F1 RT ?] Set or query the RT parameter (# is a positive integer).

RS and RT provide an alternate method of setting the ramp rate.

If the last set command for RS or RT results in both set to positive values (even if the actual setting is not changed by the command), they are used to calculate RR for actual use by the firmware and the ramp status is set to 'W'/waiting.

If ramp status is 'W'/waiting or '+'/ramping, setting both RS and RT to 0 will change the status to '-'/off (but will not change the ramp rate setting).

To Ramp the temperature:

1. equilibrate at the starting temperature;
2. set the ramp rate;
3. set a new target temperature (command class 5).

The target temperature may be above or below the current temperature; as soon as it is set, the ramp will begin, up or down, to that new target.

After reaching the target, the controller will hold at that temperature. At any time a new ramp rate and target temperature can be set to start a new ramp.

Once the ramp is completed, if you want to start another ramp (even using the same ramp rate) you must send ramp rate set command. Otherwise, setting a new target temperature later will initiate a ramp to that target temperature.

Note: The following details aid in running T-App.

The minimum settable ramp rate (using the [F1 RR S #] command) is 0.01 °C/minute. The maximum is 10 °C/min.

For higher ramp settings, the observed rate may be lower than that specified or it may be nonlinear over part of the temperature range because the maximum possible rate of heating or cooling is limited (and dependent on the ramp direction as well as on the temperature).

For compatibility with the TC 125/225/425 family of controllers, when the ramping process is completed the controller will send an [F1 TT #] response, where # will be the target temperature used to start the ramp. Depending on automatic reply settings, an [F1 RR -] and/or [F1 IS 0++C-] response may also be sent by the TC 1.

When the ramp status is '+'/On, sending a [F1 TT S #] command or a [F1 TC -] command will change the ramp status to '-'/Off.

5.11. Heat Exchanger Temperature

[F1 HT ?] What is the current temperature of the heat exchanger?

[F1 HT +3] Start periodic heat exchanger temperature reports every 3 seconds.

... [F1 HT 39] The current heat exchanger temperature is 39 °C.

[F1 HT -]	Stop periodic heat exchanger temperature reports.
[F1 HL ?]	What is the high temperature limit for the heat exchanger?
... [F1 HL 60]	The heat exchanger high temperature limit is 60 °C. In operation with temperature control on, if the HT parameter exceeds HL, the TC 1 will turn temperature control off. It will also send [F1 ER 9], [F1 TC -] and/or [F1 IS ...] (with parameter 3 = '-'/minus) depending on the current auto-reply settings (see below).

5.12. Cell Changing

Note: These commands will have an effect only for a sample holder with multiple cuvette positions.

[F2 DI]	Device initialize: move to home position.
[F2 PI]	Device initialize: move to home position and reply when done.
... [F2 DL 1]	Device is finished moving. (Original reply was OK rather than DL 1)
[F2 DL 3]	Device locate: move to position 3. (Device should be initialized prior to using this command for the first time.)
[F2 PI]	Device initialize: move to home, then to position 1 and reply when done.
... [F2 DL 1]	Device is in position 1.
[F2 PL 4]	Device locate: move to position 4 and reply when done. (Device should be initialized prior to using this command for the first time.)
... [F2 DL 4]	Device is now in position 4.
[F2 DD 400]	Set speed to 400 (acceptable range 100 – 900, with a default of 500).

5.13. Reference Cuvette

Note: These commands will have an effect only for systems with two independently-controlled sample holders.

[R1 ...]

To control and monitor the temperature and status of the reference cuvette using a Dual Temperature Controller, use any commands in classes 3-8, 10 and 11, substituting R1 for F1. There are no corresponding [R1 ...] commands for command classes 1, 2, 9 and 12.

5.14. Front Panel Controls

Note: The TC 1 can be controlled manually using the buttons and the display on the front. Once a control program has connected to the TC 1, the controller will automatically send reports to the program whenever a manual change has been made. The commands in this section allow the program to control access to front panel (FP) settings and determine if and how such front panel changes are reported.

[F1 LO (+/-)] **LockOut** (+) or do not lockout (-) the front panel. When lockout is on (+) the front panel will display "LOK" or "LLK" in the upper right corner of the display. The user will be able to use the left-arrow and right-arrow button to move between the settings displays (as if no control program was connected) but the up-arrow, down-arrow and SET buttons will not allow changes in the settings. The up-arrow and down-arrow buttons will work to move between the main "SAM", "REF" and version/ID displays.

[F1 LO ?]
 ...[F1 LO +] or ...[F1 LO -] Query the current lockout state.

[F1 LK (+/-)] **Link** (+) or unlink (-) the reference to/from the sample. This command is only available for a dual sample holder system. When link is on (+) the front panel will display "LNK" or "LKK" in the upper right corner of the display. The user will be unable to change any of the REF settings (they will be locked as described above). When the user changes a "SAM" setting using the front panel controls (such as temperature control On/Off, stirrer speed, ramp rate) the identical change will automatically be made for "REF".

[F1 LK ?]
 ...[F1 LK +] or ...[F1 LK -] Query the current link state.

T-App uses this command to turn linking off in the temperature controller and handles linking of the reference settings to the sample settings in the program (see link reference to sample).

"LKK" in the upper right corner of the display indicates that both lockout and link have been turned on by the control program. Note: Do not attempt to run a control program with the temperature controller in linked mode; that capability is not available at the current time. Only use the [F1 LK -] command to make sure the controller is not in linked mode.

[F1 FP (+/-)] Report (+) or do not report (-) changes made by use of the **F**ront **P**anel controls on the TC 1. The power-on default is to report FP changes. Setting changes are reported by [F1 TT #], [F1 SS #] and [F1 RR #] replies. The power-on default for status changes is to report by [F1 TC (+/)], [F1 SS (+/)] or [F1 RR (+/)] replies, even if the control program has not turned these status change replies. The control program cannot turn these responses off (except as described below).
 If the control program has sent [F1 IS +] or [F1 IS R+] commands to turn automatic instrument status reports on, status changes resulting from use of the front panel controls will be reported only by the instrument status reply, [F1 IS ...]. There is no query command for the FP state.

6. Elements of the T-App Software Help File

The program T-App comes with a complete Help File. Below are elements useful for getting started with T-App.

6.1. Application Startup Window

This Startup Window is displayed while the program searches for a **Temperature Controller** by checking each active COM port on the computer, beginning with the highest.



When a **Controller** is found, this startup window is replaced by the [Application Control Window](#).

If a **Controller** is not found, the startup window changes to that below. Click on the areas where the pointer changes to a hand for further information on the Application Menus and buttons,




This may result if the **Controller** is not connected to power, or if the Bluetooth or USB link between the **Controller** and the computer is not available.

Click the [Try Again button](#) (or use the [File>Restart](#) menu item) once you have corrected the problem. Click the [Exit button](#) (or use the [File>Exit](#) menu item) to terminate the program.

You may also use the [File>\(Work Offline\)](#) menu item to use the program without connecting to a sample holder in order to plot previously saved time/temperature data without being connected to a **Controller**.

6.2. Application Control Window


Click on the areas where the pointer changes to a hand for further information on the Application Menus, the Temperature Plot or the Status Panel.



The image above shows the T-App Application window as it would appear using a **TC 1 Temperature Controller** with a Series 400 thermistor probe attached (with the thermistor tip in the stirred sample). For other configurations (dual sample holders; multi-cell holders; TC 125, 225 and 425 temperature controllers) see the [Status Panel](#) topic.


6.3. Application Menus

Click on the individual menu bar items (where the pointer changes to a hand) for further information on that menu.




6.4. File Menu

Click on the individual menu item (where the pointer changes to a hand) for further information on that menu item.




6.5. Tools Menu

Click on the individual menu item (where the pointer changes to a hand) for further information on that menu item. Tools menu items for a single sample holder system.




6.6. Temperature Control

Tools>(Temperature Control) menu for a single controller, including multi-position controllers.




6.7. Temperature Ramping

Single sample holder system.




6.8. Stirrer Control

Single sample holder system.




6.9. Show Script Panel



See also the [Hide Script Panel](#) menu item.


When this menu item is clicked, the right side of the Application Control Window will open to show a vertical panel containing a text display with button controls and other information near the bottom.

The snapshot below was taken when the program was in the middle of running a script. This particular script is used to do a performance run on each sample holder (the performance run for each sample holder is included as the last page in the hardware manual).




When the Show Script Panel opens, the initial text display will be empty and the bottom controls appear as shown below left. Below right show the same area while a script is running (identical to that in the snapshot of the full window above).


The controls at the bottom of the script panel change depending on whether a delay or wait command is being executed and on whether the current command being executed is inside a loop. Click on the individual control below (where the pointer changes to a hand) for further information on that control.




NO SCRIPT RUNNING.



DELAY/WAIT COMMAND NOT INSIDE A LOOP.



CURRENT COMMAND IS INSIDE A LOOP.



DELAY/WAIT COMMAND INSIDE A LOOP.

6.10. Command Text Box

This edit box may be used to send commands to the **Temperature Controller**. The command format is illustrated by the default entry, [F1 CC P1 P2].


Replace "CC" by the command code.

Replace "P1" and "P2" by the command parameters. Most commands require only one parameter; in that case delete "P2".

Clicking the arrow at the right end of the edit box will cause a drop down list to appear showing a few valid commands.


For example, immediately after the program is started the text box will look like the figure below.

If you click on one of the items in the list, that item will appear in the box at the top and the list will disappear. You may then edit the command as needed. You may also enter a command that does not appear in the list.




6.11. Plot Menu

Click on the individual menu item (where the pointer changes to a hand) for further information on that menu item.





6.12. Data Shown

Dual with external probe



Single (or multi-sample) with external probe



6.13. Single Sample Status Panel

Click on the individual controls where the pointer changes to a hand for further information on that control.

WITHOUT PROBE ATTACHED

Sample Holder

Control Status:
 20.00 °C On
● holding
 Heat Exch: 27 °C

Target:
 20.00

Stirrer:
 On
 1500 rpm

WITH PROBE ATTACHED

Sample Holder

Control Status:
 20.00 °C On
● holding
 Heat Exch: 27 °C

Target:
 20.00

Stirrer:
 On
 1500 rpm

External Probe
 19.9 °C Avg

6.14. Temperature Control Status




When temperature control is off, the circle will be a steady gray and the label will read "off".

When temperature control is on but the sample holder temperature is not stable, the small circle will alternate slowly between green and gray and the label to its right will read "seeking".

When temperature control is on and the sample holder temperature is stable, the circle will be a steady green and the label will read "holding".

If an error condition occurs that causes temperature control to be turned off the circle and the caption to its right will alternate rapidly between red and gray.



6.15. Heat Exchanger Temperature



The temperature of the heat exchanger in the sample holder is displayed here.


The heat exchanger temperature is dependent on the temperature and the rate of flow of the water being circulated through the sample holder and on the temperature control settings.

As a warning, when the heat exchanger temperature is within 10 °C of the preset limit (typically 60 °C, the color of the label will change to orange. To correct the problem, you must reduce the temperature of the water (for example, by adding ice or cold water to the coolant container).


If the heat exchanger temperature exceeds that preset limit, temperature control will automatically be turned off to prevent damage to the system.

If temperature control does turn off, the label will turn red and an error message will display below the menu bar as illustrated below; also, the [Temperature Control Status](#) will show the error number and begin flashing red. The error message (and the flashing of the status control) can be removed at any time by clicking on the Clear button to the left of the message.

Once the problem with the heat exchanger temperature is corrected, you can restart temperature control by clicking on the [Temperature Control On/Off checkbox](#) or using the menu item [Tools>\(Temperature Control\)>On](#).



6.16. Target Temperature




The Target Temperature is the Sample Holder Temperature that you want to use. If Temperature Control is On, the **Controller** will either be driving the holder temperature to the target or maintaining the holder temperature at the target.

The number displayed is the current Target Temperature. When the Change button to the right of the number is clicked a small parameter entry dialog will appear where you can change the Target Temperature setting.

The [Tools>\(Temperature Control\)>\(Set Target\)](#) menu item may also be used to change the Target Temperature.

6.17. Magnetic Stirrer Controls




The checkbox may be used to turn the magnetic stirrer on (checked) or off (unchecked).

Click on the Change button to reset the speed of the stirrer. The current speed setting is shown to the left of the button. The submenus of the [Tools>\(Stirrer Control\)](#) menu item may also be used to change the stirrer properties.

6.18. External Probe Status

The External Probe Status will be displayed only if a Series-400 or Series-500 thermistor probe is connected to the 1/4 inch phone jack in the back of the sample holder. This probe input is intended to provide a means to directly measure the temperature of the sample in the cuvette.



The number to the left of the Avg checkbox is always the most recent temperature from the probe. If a probe temperature is not yet available, the digits will be replaced by dashes.


The left image shows the display for when the Avg box is not checked..

When the Avg checkbox is checked (by clicking on it), all controller models send probe readings with two decimal digits. The numbers in parentheses beneath the checkbox (middle image) list the mean value of the most recent 20 temperatures for the probe followed by the average deviation from that mean. The mean value should provide a more reproducible sample temperature if it has been stable long enough for 20 temperatures from the stable period to have been acquired.

If 20 temperatures are not yet available, the mean and average deviation are replaced by dashes (image on the right).

6.19. Plot Axis Label Manipulation


The label on the vertical axis can be changed by clicking on the small black rectangle near the upper-left corner of the plot; this will make a text box appear for changing the label. After new text has been entered, hitting the Tab or Return key will cause the text box to close and the modified axis label to appear on the plot.



BEFORE CLICKING ON THE BLACK RECTANGLE.

AFTER CLICKING ON THE BLACK RECTANGLE.

The label on the horizontal axis can be changed by clicking on the small black rectangle near the lower-right corner of the plot; this will make a text box appear for changing the label. After new text has been entered, hitting the Tab or Return key will cause the text box to close and the modified axis label to appear on the plot.




BEFORE CLICKING ON THE BLACK RECTANGLE.

AFTER CLICKING ON THE BLACK RECTANGLE

6.20. Plot of Data Collected

This plot shows an example of temperature as a function of time data that would result from using the [Example Script](#).



black - temperature of the sample holder. blue - temperature of the heat exchanger.

dotted black - temperature of the [external probe](#) (with the thermistor tip in the liquid sample).

Individual time/temperature points are shown for the external probe data to illustrate special methods of reporting probe temperatures used in the Example Script. Normally the probe data are shown as a red line.

6.21. Scripts: Program Script Commands

The Program Script Commands are used to control the behavior of a script and are not sent to the **QNW Serial Controller**.

Delay and Wait commands

Progress of the Delay and Wait commands is tracked in the [Delay/Wait Countdown Display](#) at the bottom of the [Script Panel](#).

[*D #] - Delay # INTERVALS before running the next command. Example: [*D 120] will cause a delay of 120 of the INTERVAL units as defined in the first line of the script (see the [Example Script](#) which sets the INTERVAL to 0.01 minutes).

While the delay is active, a countdown of the number of script INTERVALs remaining is included in the **Countdown Display** (see **Window View 2** in the **Script Panel**).

[*WRP>=#] **[*WRP<=#]** These commands are included to be compatible with older scripts. T-App now interprets them as **[*WCT>=#]** and **[*WCT<=#]**

[*WCT>=#] - Wait until the current (holder) temperature exceeds or equals a particular value (# as an integer).

[*WCT<=#] - Wait until the current temperature decreases to or below a particular value (# as an integer).

The current temperature is the temperature of the metal body of the sample cuvette holder.

[*WPT>=#] - Wait until the probe temperature exceeds or equals a particular value (# as an integer).

[*WPT<=#] - Wait until the probe temperature decreases to or below a particular value (# as an integer).

The probe temperature is the reading of the standard 400 series or 500 series probe (usually the probe is inserted in the sample cuvette).

[*WRT>=#] - Wait until the reference temperature exceeds or equals a particular value (# as an integer).

[*WRT<=#] - Wait until the reference temperature decreases to or below a particular value (# as an integer).

The reference temperature is the temperature of the metal body of the reference cuvette holder (dual controller only).

[*WT 1000 2] wait for stable T (or 20 min max)

[*WT #] - This command will be interpreted as a **[*WT 1000 1]** command to maintain compatibility with older script files.

[*WT (#1) (#2)] - This program command will cause the script to Wait until the sample holder Temperature is stable before continuing to the next script command in the list. Temperature stability is determined by an algorithm built into the **QNW Temperature Controller**. The parameter (#1) determines how often the program sends a status query to the controller (**[F1 IS ?]**, see the [Controller Serial Command List](#)). Each time this query is received, the controller returns a status reply which includes information as to whether the temperature is still changing or has met the criteria for being stable (controlled). The parameter (#2) specifies the maximum number of status queries that will be sent before the wait is canceled.

The time between status queries is the product of the parameter (#1) and the script INTERVAL as defined in the second line of the script (see the [Example Script](#)).

While the wait is active, the number of times the controller has been queried is displayed followed by a countdown to the next query.

Note: if the command **[F1 IS +]** has been sent to the controller, the the instrument status reply will be sent by the **QNW Temperature Controller** automatically when the sample holder temperature becomes stable. In this case the command form **[*WT (#1) 1]** is recommended where (#1) is chosen to give a 10 minute period (for example, if INTERVAL IS 0.6 seconds or 0.01 minutes, then (#1) should be set to 1000). Typically, stable sample holder temperature is attained within 10 min of setting a new target temperature.

[*WD #] - This command is no longer accepted because it does not work in Windows XP or later.

Beep Control commands

[*BCT +] or **[*BCT-]** - Turn on(+) or off(-) beeping by the computer whenever a sample holder (current) temperature report is received from the controller.

[*BPT +] or **[*BPT-]** - Turn on(+) or off(-) beeping by the computer whenever a probe temperature report is received from the controller.

[*BRT +] or **[*BRT-]** - Turn on(+) or off(-) beeping by the computer whenever a reference holder temperature report is received from the controller.

Warning Control commands

[*E+] or **[*E-]** - The commands are no longer supported because they are not needed for **T-App** version 2.21.

Listing Control commands

[*LIS +] or **[*LIS -]** - Enable (+) or Disable (-) listing of Instrument Status replies during running of a script.

[*LER +] or **[*LER -]** - Enable (+) or Disable (-) listing of Error Message replies during running of a script.

[*LCT +] or **[*LCT -]** - Enable (+) or Disable (-) listing of Current Temperature replies during running of a script.

[*LPT +] or **[*LPT -]** - Enable (+) or Disable (-) listing of Probe Temperature replies during running of a script.

[*LRT +] or **[*LRT -]** - Enable (+) or Disable (-) listing of Reference Temperature replies during running of a script.

These ten program commands can be used to control whether the indicated information, sent from the Serial Controller to the program, are to be listed (displayed) in the Script Panel or the Command Panel while a script is running. Disabling the listings may be useful to make progress of the script clearer since, under some conditions, large numbers of Instrument Status replies and Temperatures replies may be generated while the script is running. Disabling affects only the listing of these various replies -- the replies will still have their normal effects in the Main Window.

Note: some earlier versions of **T-App** defaulted to listing temperature replies in the Command/Script panel. **T-App** version 1.51 defaults to not listing them.

Miscellaneous commands

[*R] - This program command is used only at the end of a script. It will cause the script to be repeated, starting at the beginning. The script will then repeat forever or until the [Cancel Script](#) button is clicked.

[*P] - If the plot window is displayed while a script is running, the data in the plot may not be automatically updated unless a delay (*D) or wait (*W) command is being processed. This program command will cause the plot window to update (replot to show all data collected to that point).

[*CTD] - This command will clear the time and temperature data from the [Time/Temperature Plot](#) of the [Application Control Window](#). In addition, the time parameter in the displays will be restarted at zero and listing of time/temperature data in the display windows will be started. Whatever display interval was in effect at the time this command is executed is retained (set by a script command or by clicking on the (Plot>Set Interval) menu item).

[*MSG + message] or **[*MSG - message]** - This command will stop execution of the script and present a dialog box containing the second command parameter ("message"); change this parameter to the wording you want the dialog box to display. The first parameter (+ or -) determines whether the computer beeps (+) or not (-) while the dialog box is visible. Clicking on the "OK" button in the dialog box closes it and execution of the script continues.

[*LS #] and **[*LE]** - The **[*LS #]** command defines the start of a loop (a repeat of a series of commands) within the script; the parameter '#' is an integer specifying how many times the loop is to be run. The **[*LE]** command defines the end of the series of commands included in the loop. A script may contain more than one loop. Also, smaller loops can occur within larger loops.

[*WPL] -- When the positioning commands **[F2 PL #]** = go to position #

or **[F2 PI]** = go to home and back to the last position

are sent to the **Temperature Controller**, it automatically sends the reply [F2 DL #] when the positioning process is completed. The [*WPL] command waits for this reply before continuing to the next command in the script.

[*PL+] -- This command will cause the sample position to change to the next higher number. If the current position is the highest number position, it will move to position 1.

[*PL-] -- This command will cause the sample position to change to the next lower number. If the current position is 1, it will move to the highest number position.

[*TT+#] or **[*TT-#]** -- Increments or decrements the sample holder target temperature.

[*RT+#] or **[*RT-#]** -- Increments or decrements the reference holder target temperature.